



GIMA

PROFESSIONAL MEDICAL PRODUCTS

KIT 5010E MATERASSO A PRESSIONE ALTERNATA+POMPA
KIT 5010E ALTERNATING PRESSURE MATTRESS + PUMP

All serious accidents concerning the medical device supplied by us must be reported to the manufacturer and competent authority of the member state where your registered office is located.

INSTRUCTIONS

The Alternating Pressure Mattress consists of two parts: the mattress and the pump. The pump utilizes a small compressor to remain quiet and energy efficient. The control panel is simple and very user friendly. The mattress relieves pressure by sequentially deflating and inflating alternate air cells on approximately 12 minutes timed intervals. It is widely recognized that constant pressure to a bony prominence is the leading cause of skin breakdown. The continuous movement supplied by the unit lessens the areas of constant pressure and enhances circulation.

OPERATION

1. Place the mattress on the bed frame with the hose end at the foot section of the bed frame.
2. Using the integrated hooks, securely hang the pump on the bed end at the foot end or place on a smooth flat surface.
3. Connect the air hoses from the mattress to the pump.
4. Plug the pump into a wall outlet. Be sure the power cord is safely away from possible hazards.
5. Turn on the power switch on the control panel of the pump. The pump will begin to inflate the mattress.
6. After inflation, adjust the mattress using the dial on the pump.

CLEANING INSTRUCTION

Mattress

Clean the mattress regularly with neutral detergent or alcohol. Heating or steam sterilization is not available.

Pump

Clean the pump regularly with neutral detergent or alcohol. Do not open the pump housing – risk of electric shock. Do not get the pump wet or submerge it in any liquid.

CAUTION

1. Do not smoke on or near the pump.
2. Keep the pump away from heated surfaces.
3. Explosion risk if used in the presence of flammable anesthetics.
4. Replace fuse as marked: T1A 250V. Power Input: ~ 220 V - 50 Hz - 0,1 A

PRODUCT SPECIFICATIONS

WEIGHT CAPACITY		ELECTRICAL	
Maximum patient weight	331lbs (150kg)	Operating power:	~ 220 V 50 Hz 0,1 A max.
		Fuses::	T1A 250 V
ENVIRONMENTAL CONDITIONS			
OPERATING CONDITIONS		STORAGE AND SHIPPING CONDITIONS	
Ambient Temperature:	10°C to 40°C (50°F to 104°F)	Ambient Temperature:	-18°C to +43°C (0°F to 110°F)
Relative Humidity:	10% to 75%	Relative Humidity:	10% to 95%

	Keep in a cool, dry place		Keep away from sunlight		Temperature limit
	WEEE disposal		Manufacturer		Humidity limit
	Follow instructions for use		Type BF applied part		Caution: read instructions (warnings) carefully
	Product code		Lot number		Serial number
	Class II applied		Medical Device compliant with Regulation (EU) 2017/745		Medical Device
	Imported by		Authorized representative in the European community		Date of manufacture
	Unique Device Identifier				

Disposal: The product must not be disposed of along with other domestic waste. The users must dispose of this equipment by bringing it to a specific recycling point for electric and electronic equipment.

GIMA WARRANTY TERMS

The Gima 12-month standard B2B warranty applies.

Guangdong Yuehua Medical Instrument Factory Co., Ltd.
Rongsheng Science and Technology Zone,
Daxue Road, Shantou, Guangdong, China
Made in China

Eunitor GmbH
Kennedydamm 5,
40476 Düsseldorf, Germany

QDC-5010E+P3000N2EB

Gima 28575

Gima S.p.A.
Via Marconi, 1 - 20060 Gessate (MI) Italy
gima@gimaitaly.com - export@gimaitaly.com
www.gimaitaly.com
Made in China

