



# Piccolo Basic – Plus – Speech

PC & iPad CONTROLLED AUDIOMETER

## DESCRIPTION

Piccolo is an audiometer controllable either through any Windows computer or through an iPad.

Piccolo is available in three different models:

- Basic* : only PTA – AC
- Plus* : only PTA – AC / BC
- Speech* : PTA & Speech – AC / BC

When controlled through the computer (using Inventis Daisy or the dedicated Noah module), Piccolo is connected to it via USB.

To control it through the iPad via Bluetooth 4 connection, Piccolo needs to be purchased with the *Aero* option.

## CLASSIFICATION

Basic: EN 60645-1 / ANSI S3.6: Type 4  
 Plus / Speech: EN 60645-1 / ANSI S3.6: Type 3

## AVAILABLE SIGNALS

Stimulus: pure tone, warble tone  
 Masking: NBN, WN, SN (Piccolo Speech only)

Only *Piccolo Speech*:

2 external inputs for speech audiometry  
 MIC input for live speech audiometry

## SIGNALS SPECIFICATION

Attenuator step: 5 dB  
 Presentation: Continuous, Pulsed (0.5, 1 and 2 Hz)  
 Warble: 5 Hz sin wave modulating signal

## AVAILABLE OUTPUTS AND TRANSDUCERS

AC: TDH-39, ER-3A / ER-5A insert earphones  
 BC: B-71 bone vibrator (Piccolo Plus / Speech only)

## AVAILABLE TESTS

- Pure Tone audiometry
- Auto threshold (modified Hughson-Westlake)
- Speech audiometry (Piccolo Speech only)

## PURE TONE: FREQUENCIES AND MAXIMUM LEVELS (dB HL)\*

Freq. (Hz)	AC TDH-39	AC ER-3A	AC ER-5A	BC
125	80	90	90	-
250	100	105	100	45
500	110	110	110	65
750	115	115	120	70
1.000	120	120	120	75
1.500	120	120	120	80
2.000	120	120	115	80
3.000	120	120	115	75
4.000	120	110	110	75
6.000	105	95	100	55
8.000	95	90	90	50

(\*The maximum intensities are obtainable only if the audiometer is powered with the mains adapter.

## SPEECH AUDIOMETRY: MAXIMUM LEVELS (dB HL)\*

AC TDH-39	AC ER-3A	AC ER-5A	BC
100	100	100	55

(\* Only available for the Speech version.

## PATIENT – OPERATOR COMMUNICATION

Talk over: external microphone included  
 Patient response trigger

## CALIBRATION

Validity: 12 months  
 All the parameters set through the device software

## COMPUTER INTERFACE

Connection: USB (driverless)  
 Compatible software: - Inventis Daisy with Maestro module  
 - Noah with “Maestro for Noah” module

## INTERFACING WITH THE iPad

Through Bluetooth 4 connection (needs *Aero* option).  
 Requires iPad 3 or successive  
 Dedicated app freely downloadable from the Apple App Store

## POWER SUPPLY

Maximum consumption: 9 Watts  
 Power supply: 6V, 2A cont., through an external medical grade 100-240 Vac 50/60 Hz power supply

## MECHANICS

Size (WxDxH): 16 x 16 x 3 cm / 6.3 x 6.3 x 1.2 in  
 Weight: 1.8 Kg / 4.0 lbs

## FREIGHT PACKING

Size (WxDxH): 40 x 30 x 18 cm / 15.8 x 11.8 x 7.1 in  
 Gross weight: 2.7 Kg / 5.9 lbs

## APPLICABLE STANDARDS

Pure tone audiometry: EN 60645-1, Type 3 (Piccolo Plus / Speech)  
 EN 60645-1, Type 4 (Piccolo Basic)  
 Calibration: EN ISO 389-1 (TDH 39), EN ISO 389-2 (ER-3A and ER-5A), EN ISO 389-3 (B71)  
 Electrical safety: EN 60601-1, Class I type BF  
 EMC: EN 60601-1-2

## CE CERTIFICATE

93/42/EEC classification : Class IIa  
 Classification rule (Annex IX, 93/42/EEC): 10  
 Notified body: TÜV SÜD Product Service GmbH (0123)  
 Number of CE certificate: G1 12 09 65346 006

## PRODUCT CODES

Code 10327 - Piccolo model Basic  
 Code 10328 - Piccolo model Basic Aero  
 Code 10329 - Piccolo model Plus  
 Code 10330 - Piccolo model Plus Aero  
 Code 10331 - Piccolo model Speech  
 Code 10332 - Piccolo model Speech Aero

**INCLUDED PARTS**

- TDH-39 supra-aural headphones
- B71 bone vibrator (Piccolo Plus / Speech only)
- Patient response switch
- Clip-on microphone for patient-to-operator communication
- Medical grade power supply
- USB connection cable
- Jack 3.5mm connection cable (Piccolo Speech only)
- User manual
- Inventis Software Suite CD
- Noah Maestro module
- Daisy Maestro module

**OPTIONAL ACCESSORIES (with order code)**

- 10166: ER-3A insert earphones
- 10172: ER-5A insert earphones
- 10179: Amplivox Audiocups noise excluding enclosures for

Piccolo is developed by:

**INVENTIS s.r.l.**

CORSO STATI UNITI, 1/3  
35127 PADOVA – ITALY  
PHONE: +39.049.8962 844  
FAX: +39.049.8966 343  
info@inventis.it  
www.inventis.it

Follow us on Twitter  
<http://twitter.com/inventissrl>



The Inventis Quality System complies with ISO 9001 and ISO 13485 standards.

Inventis® is a registered trademark of Inventis s.r.l.