

LIFE ♥ SENSE®



**WIDESCREEN™**  
CAPNOGRAPHY & OXIMETRY

The LifeSense® multi-parameter monitor is ideal for continuous monitoring in critical care environments. Equipped with NONIN PureSAT® SpO<sub>2</sub> and MedAir™ EtCO<sub>2</sub> technologies, LifeSense is a proven performer for ET-tube placement verifications, waveform trend monitoring, detecting breathing irregularities, gauging the efficacy of CPR and procedural sedation monitoring. Its sidestream design allows use with intubated and non-intubated patients.

**Features:**

- » Widescreen touch panel display
- » NONIN PureSAT SpO<sub>2</sub> technology
- » MedAir EtCO<sub>2</sub> technology
- » Numerical pulse rate and respiratory rate display
- » Numerical EtCO<sub>2</sub> and SpO<sub>2</sub> display
- » 4-hour waveform trending
- » Capnograph and plethysmograph display
- » Backlit LCD display
- » Audible and visual alarms
- » Data output via RS-232 port

**Options:**

- » Full line of infant to adult SpO<sub>2</sub> sensors (reusable and disposable)
- » Full range of infant to adult sampling lines
- » Carrying bag
- » Adjustable mounting clamp
- » TrendSense™ memory module



Infant to Adult CO<sub>2</sub> Sampling Cannulas



T-connectors for Intubated monitoring



Wide range of PureLight® SpO<sub>2</sub> Sensors

# LIFE SENSE®

## Power Data

Battery charger:	120 VAC, 60 Hz for USA / Canada
Power consumption:	3.6 W at battery operation 12 W with battery charger
Input:	12 VDC, 800mA

## Physical Data

Dimensions:	200 (W) x 135 (H) x 50 (D) mm (7.9 x 5.3 x 2 inches)
Weight:	800 grams (1.8 lbs)

## Operation

Working temperature:	-5° to +40°C (23° to 104°F)
Humidity:	10 – 95% (non-condensing)
Atmospheric pressure:	860 – 1060 hPa

## Storage

Storage temperature:	-20° to +50°C (-4° to +122°F)
Humidity:	10 – 90% (non-condensing)
Atmospheric pressure:	up to 4 atmosphere (110 - 4050 hPa)

## Pump

Pump flow:	75 ml/min
Flow accuracy:	±15ml/min

## Battery Data

Type:	Lithium Ion internal battery, rechargeable
Battery capacity:	8 hours approximately
Charging time:	17 hours from empty to full or 2 hours for every hour to run on battery

## Capnography Measuring Data

Respiration range:	3 – 80 respirations/min
Update frequency:	Once every breath (No breath after 25 seconds)
Respiration accuracy:	3 – 50 respirations/min ± 2 51 – 80 respirations/min ± 5
EtCO <sub>2</sub> /CO <sub>2</sub> range:	0 – 9.9 kPa, or 0 – 99 mmHg
EtCO <sub>2</sub> /CO <sub>2</sub> accuracy:	±0.2 kPa / ±2 mmHg, +6% of reading (EtCO <sub>2</sub> /CO <sub>2</sub> reading reaches its steady state accuracy 10 minutes after power up)
Update frequency:	Once every breath (No breath)
Trend length:	4 hours

## Pulse Oximeter Specifications


<b>Displayed Oxygen Saturation Range (SpO<sub>2</sub>)</b>	0 to 100%
<b>Displayed Pulse Rate Range</b>	18 to 255 beats per minute (BPM)
<b>Measurement Wavelengths***</b>	Red: 660 nanometers @ 0.8 mW maximum average Infrared: 910 nanometers @ 1.2 mW maximum average
<b>Using NONIN PureLight® Sensors</b>	
<b>Saturation Accuracy (Arms) **</b>	70 - 100%

## Sensor Specifications

Reusable Finger Clip (8000AA-2)	
<b>Adult / Pediatric</b>	± 2 digits*

Specifications are subject to change without notice

Authorized Distributor



\*Specifications based on 8000AA PureLight® finger clip sensor

\*\*±1 Arms represents approximately 68% of measurements

\*\*\*This information is especially useful for clinicians performing photodynamic therapy

Nonin Medical, Inc.  
13700 1st Avenue North, Plymouth, MN 55441-5443, USA  
Phone: +1 763.553.9968 | Toll Free: 800.356.8874 | Fax: 763.553.7807 | Email: info@nonin.com

www.nonin.com

Nonin®, PureSAT®, PureLight® and TrendSense™ are trademarks of Nonin Medical, Inc.  
LifeSense® and MedAir™ are trademarks of MedAir AB.



©2007 Nonin Medical, Inc. P/N 6025-000-01